



Certificate of Analysis

Product:
Percoll™ PLUS

Code Number:
17544502

Lot No: 10326907

Test/Characteristic:	Limit:	Result:
Density at 20°C g/mL	1.125 - 1.135	1.131
Viscosity cP	max. 15	6
Osmolality mOsmol/kg	max. 30	14
Carbon content in dry residue %	4.0 - 5.5	4.9
pH	8.9 - 9.9	9.3
Endotoxin activity EU/mL	less than 1	less than 1

**Manufactured in compliance with our ISO 9001 certified quality management system.
No animal derived material is used in raw materials or in the manufacture of this product.**

Approval Date (Year-Month-Day): 2022-06-30

Expiry Date (Year-Month): 2027-05

Manufacturing Date (Year-Month): 2022-05

Cytiva and the Drop logo are trademarks of Global Life Sciences IP Holdco LLC or an affiliate

Cytiva Sweden AB
Björkgatan 30
SE-751 84 Uppsala, Sweden
T + 46 (0)18 612 00 00
Cytiva.com
Reg.No. SE 55 61 08 1919 01

Tests and limits according to AS 28903926 Ed. AF

Issued (Year-Month-day): 2022-06-30 by Quality Assurance

Vällfors, Christofer

This document has been electronically produced and is valid without a signature.

DOC1103901 / 7