



Certificate of Analysis

Product:
Percoll™ PLUS

Code Number:
17544502

Lot No: 10290196

Test/Characteristic:	Limits:	Results:
Density g/mL	1.125 - 1.135	1.131
Osmolality mOsmol/kg	max. 30	17
Viscosity cP	max. 15	6
Carbon content in dry residue %	4.0 - 5.5	4.8
pH	8.9 - 9.9	9.3
Endotoxin activity EU/ml	less than 1	less than 1

Manufactured in compliance with our ISO 9001 certified quality management system.
No animal derived material is used in raw materials or in the manufacture of this product.

Approval Date (Year-Month-Day): 2020-04-21

Expiry Date (Year-Month): 2025-03

Manufacturing Date (Year-Month): 2020-03

GE Healthcare Bio-Sciences AB
Björkgatan 30
SE-751 84 Uppsala, Sweden
T + 46 (0)18 612 00 00
www.gehealthcare.com
Reg.No. SE 55 61 08 1919 01

Tests and limits according to AS 28-9039-26 Ed. AE

Issued (Year-Month-day): 2020-04-21 by Quality Assurance

Medin, Britta

This document has been electronically produced and is valid without a signature.