

# Certificate of Analysis

## LV-MAX™ Lentiviral Packaging Mix

Product No. A43237

Lot No. 2509813A

Date of Manufacture 28-Jul-2022

---

### Restriction Enzyme Analysis

LV-MAX™ Lentiviral Packaging Mix consists of a mixture of the pLP1, pLP2, and pLP/VSVG plasmids. The mixture of the three plasmids is qualified by restriction analysis. Restriction digests must demonstrate the correct banding pattern when electrophoresed on an agarose gel. The tables below list the restriction enzymes and the expected fragments.

#### *LV-MAX™ Lentiviral Packaging Mix*

<u>Restriction Enzyme</u>	<u>Expected Fragments (bp)</u>
<i>Afl II / Age I</i>	7630, 5821, 4180, 1259

### Purity

To ensure purity, the absorbance at 260 nm ( $A_{260}$ ) and the absorbance at 280 nm ( $A_{280}$ ) of the vector preparations are measured by spectrophotometry. The  $A_{260}/A_{280}$  ratio must be between 1.8 - 2.0.

### Results

Meets all specifications

---

For Research Use Only. Not for use in diagnostic procedures.

Thermo Fisher Scientific  
Life Sciences Solutions  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
<https://www.thermofisher.com>  
For inquiries, contact us at [cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)



Chevohn Joseph  
Director, Quality  
Issued on 19-Jan-2023